



إعلان

**مناقصة 22/2018**

**لتوريد جهاز ألتراساوند وأجهزة تعقيم بدعم من جمعية التضامن المغربي**

فعلى الشركات الراغبة بدخول العطاء مراجعة اتحاد لجان العمل الصحي الكائن في غزة - النصر - أبراج المقوسي للحصول على العطاء والأصناف المطلوبة ابتداءً من يوم السبت الموافق 2018/12/8 وحتى نهاية عمل يوم الثلاثاء الموافق 2018/12/11. مع العلم بأن آخر موعد لتسليم عروض الأسعار هو نهاية عمل يوم الأحد الموافق 2018/12/16. علماً بأن عرض السعر يسلم مقابل ( 100 دولار ) غير مستردة

ملاحظة : تم إدراج أوراق المناقصة على الموقع الإلكتروني للاتحاد على العنوان التالي: [www.uhwc.ps](http://www.uhwc.ps)

للاستفسار عن المناقصة الرجاء الاتصال بـ :السيدة/ نهال مهنا - قسم المشتريات على هاتف رقم **2895987**

**مواعيد العمل في الاتحاد من الساعة الثامنة صباحاً وحتى الساعة الثانية من بعد الظهر. مع التأكيد على عدم إمكانية بيع أي عطاء بعد التاريخ المحدد.**

رئيس قسم المشتريات

نهال مهنا



## الشروط الخاصة:

1. يشمل العرض على توريد جهاز ألتراساوند وأجهزة تعقيم لصالح اتحاد لجان العمل الصحي حسب البنود الموضحة في جدول الكميات والأسعار.
2. الأسعار بالدولار و شاملة قيمة الضريبة المضافة .
3. تقوم الشركة بتقديم كفالة دخول عطاء تقدر ب 10 % من إجمالي قيمة عرض السعر.
4. تقوم الشركة بتقديم كفالة حسن تنفيذ تقدر ب 10% من إجمالي قيمة الاصناف التي تمت الترسية.
5. تقوم الشركة بتقديم كفاله ضمانه بنكية بقيمة 10 % من إجمالي أمر التوريد سارية لمدة عامان من تاريخ الاستلام.
6. يجب أن تكون الأسعار سارية المفعول لمدة 90 يوماً من تاريخ تقديم عرض السعر.
7. يلتزم المورد بتقديم شهادة منشأ وسيرته الذاتية ترفق مع عرض السعر .
8. ضرورة ارسال كتالوجات أصلية مع عرض السعر .
9. مدة التوريد خلال شهرين من تاريخ امر التوريد.
10. توريد جميع البنود المطلوبة إلى المخازن الرئيسية لاتحاد لجان العمل الصحي .
11. الاتحاد غير ملزم بقبول أقل الأسعار ودون إبداء الأسباب.
12. يجب تعبئة جميع البيانات الخاصة بالبند المراد تقديم عرض سعر له حسب ما هو موضح في جدول الكميات والأسعار .
13. طريقة الدفع: 100 % من قيمة أمر التوريد بشيك بعد أسبوعين من استلام جميع البنود الموردة.
14. رسوم دخول العطاء (100 دولار) غير مستردة.
15. آخر موعد لتسليم عروض الأسعار نهاية عمل يوم السبت الموافق 2018/12/15.



بيانات مقدم العطاء

اسم الشركة / المستودع : \_\_\_\_\_  
رقم المشتغل المرخص : \_\_\_\_\_  
المدير المسئول : \_\_\_\_\_  
العنوان : \_\_\_\_\_  
الهاتف : \_\_\_\_\_  
الفاكس : \_\_\_\_\_  
البريد الإلكتروني: \_\_\_\_\_  
رقم الحساب البنكي: \_\_\_\_\_  
البنك والفرع: \_\_\_\_\_

ممثل الشركة	
العنوان	
الهاتف	
الفاكس	
الجوال	
البريد الإلكتروني	

تفاصيل الكفالة البنكية	
اسم البنك / الفرع	
رقم الكفالة البنكية	
قيمة الكفالة البنكية	
تاريخ صدور الكفالة	
صلاحية الكفالة البنكية	



التاريخ: 2018/12/6

السادة/ \_\_\_\_\_ المحترمين

تحية طيبة وبعد،،،

الرجاء تزويدنا بعرض سعر للوارد في الجداول أدناه وحسب المواصفات والكميات الموضحة، علماً بأن آخر موعد لاستلام عروض الأسعار هو نهاية عمل يوم السبت الموافق 2018/12/15. على أن يتم التوريد خلال شهرين من تاريخ إصدار أمر التوريد كحد أقصى، على أن تكون الأسعار بالدولار شاملة لقيمة الضريبة المضافة.

No.	Item	Qty.	Unit Price (\$)	Total Price (\$)
1	Ultrasound device	1		
2	Autoclave device	3		
Total \$				

وتفضلوا بقبول فائق الاحترام والتقدير،،،

نهال مهنا

رئيس قسم المشتريات



## ULTRASOUND

- ❖ MOBILE COLOUR DOPPLER ultrasound (dedicated for OBS/GYN).
- ❖ Brand name & Well know manufacture (USA, JAPAN, EUR, Korea).
- ❖ System should have EC or CE & FDA (US) approval.
- ❖ Display mode: B, B/M, M, PWD, Color Doppler, Power Doppler Duplex, and Triplex.
- ❖ The system should have extensive calculation software package for General imaging, vascular studies and Obs/Gyn.
- ❖ Measurement function: Distance; Area, Circumference, Weight, Heart rate, CDH (HIP joint angle). Compatible with Obs/Gyn measurements.
- ❖ Color Doppler & Power Doppler
- ❖ CD/DVD-R.W. /USB facility should be available.
- ❖ Built-in hard disk drive (H.D.D) of 500 GB at least
- ❖ Separate Video output.
- ❖ LAN Port / Ethernet port (or Wireless LAN Support).
- ❖ Direct connectivity to USB Laser Printer/Thermal Printer for printing images & reports.
- ❖ Time Gain Compensation with easy gain and frequency adjustment
- ❖ THI and TDI
- ❖ Body mark and PROBE mark facility.
- ❖ Character display facility.
- ❖ Medical grade monitors 19" at least.
- ❖ PROBE selector with easy gain and frequency adjustment
- ❖ 2probe ports at least & Probes with heavy duty cable.
- ❖ High resolution transducer at least 128 elements.
- ❖ 256 (8 bits) discrete gray levels
- ❖ Complete set of stander accessories: Probe hanger, Probe holder, Dust cover.
- ❖ Integrated Mobile trolley with shelves from the same manufacturer as the Equipment complementary.
- ❖ Gel Warmer & 5L of ultrasound gel.



- ❖ Should have active online UPS for 10 minutes or more.
- ❖ Power Supply: AC 230 ±10, 50 Hz.
- ❖ B/W thermal printer is included.

**Complete with all standard accessories and ready for use.**

**Probes:**

- ❖ Convex probe multi frequency includes the range 2-6MHz.
- ❖ Endo cavity probe multi frequency includes the range 4-9MHz.

**Options priced separately**

- ❖ Linear Probe Multi frequency 4-12 MHz
- ❖ Biopsy guide for Endo cavity probe ( reusable).
- ❖ External 32" monitor full HD.

**Warranty:**

- ☒ 3 years warranty including spare parts, preventive maintenance.
- ☒ The Supplier must provide a letter from the manufacturer of the system guaranteeing availability of spare parts for at least 10 years.
- ☒ System Software Should be delivered with the unit.
- ☒ Complete unit will be delivered, installation, operation with all standard accessories.
- ☒ Operation and Service Manuals in English must be included in Soft and hard copies.



## Autoclave 40 L

- Electrical control full automatic cycle .
- Horizontal design .
- Capacity: 40 Liter stainless steel chamber .
- Programs: all program parameters can be altered and stored .
- Power: 220VAC/50 Hz .
- Audible and visual alert upon cycle interruption and completion .
- Double safety locking device .
- Outer and inner chamber made of stainless steel.not less 316L
- Overheating protection .
- At least wo trays are included.not less than 316L
- Water level indicator with automatic low water level cut off device .
- Pressure and temperature controlled and monitored .
- Pressure and temperature alarms .
- Design Temperature 140 C
- LCD display .
- Pressure 1.2-2.1 KG/CM2
- Auto dray function
- Safety valve is included .
- Supplied with standard accessories .